

# EQA testing in Belgium: flow cytometry

A. Gothot, MD, PhD

CHU Liège – Unilab Lg



	CD34 numeration	Lymphocyte subsets
#surveys	2-3	3
#samples/survey	2	2
type	Stabilized blood (if possible fresh cord blood)	Fresh K <sub>2</sub> EDTA blood
Participants	20	49

	FC/21357	
	Median	SD
% CD34+ cells within total WBC	0.170	0.022
Absolute CD34+ cell count (cells/ $\mu$ L)	10.0	1.3

FC/21559	Median	SD	CV,%	N
WBC 10E9/L	6.80	0.16	2.4	42
Lympho% haematology analyser	35.0	2.0	5.7	41
Lympho% flow cytometer	35.2	2.8	8.0	39
CD3 %	77.4	3.0	3.9	45
CD3 10E9/L	1.837	0.190	10.4	44
CD4 %	53.6	3.7	6.9	45
CD4 10E9/L	1.261	0.127	10.0	44
CD8 %	19.5	1.1	5.7	45
CD8 10E9/L	0.465	0.035	7.5	44
CD19 %	9.0	1.3	14.8	45
CD19 10E9/L	0.208	0.044	21.3	44
NKcells %	13.0	2.5	19.4	45
NKcells 10E9/L	0.309	0.072	23.2	44
Kappa % B lymphocytes	59.8	1.9	3.2	38
Lambda % B lymphocytes	39.9	2.1	5.4	38
Kappa/lambda	1.51	0.13	8.8	38
Sum K+L % B lymphocytes	99.9	0.6	0.6	38
Lymphosum %	99.4	1.0	1.0	45

# Clinical context

- Lymphocyte subsets
  - Work-up of lymphocytosis – lymphoproliferative disease screening
  - Screening/monitoring of immune deficiencies
    - Congenital ID (+ T/B cell subclasses)
    - Acquired ID: CD4+ cells in HIV infection
- CD34 cell count
  - Monitoring of stem cell mobilisation/collection procedures (peripheral blood)
  - Cell therapy products QC: apheresis, cord blood

# Lymphocyte subsets

Annual report 2024

# Double vs single platform

## Double platform

- **Flow Cytometer:**

% target cells\* relative to total CD45+ leucocytes



- **Hematology analyser:**

#leucocytes/ $\mu$ L

## Single platform


- **Flow Cytometer:** % target cells\* relative to beads



- Beads with defined concentration in #beads/ $\mu$ L

\* CD34+ cells or lymphocytes

# Single platform lymphocyte numeration



# Equipment

- 12% single platform, 88% double platform

Haematology analyser	Number of participants
Sysmex XN 1000/ XN 2000/ XN 3000/ XN 9000	39
Beckman Coulter UniCel DxH 800 / DxH 900	5
Siemens Advia 2120	1
Abbott Cell-Dyn Ruby	1
Not mentioned	3

<i>Flow cytometer</i>	Number of participants
BD Biosciences FACSLyric	20
Beckman Coulter Navios	10
BD Biosciences FACSCanto II	9
Beckman Coulter DxFLEx	5
Beckman Coulter AQUIOS CL	2
Beckman Coulter Cytomics FC 500	1
BD Biosciences FACSVia	1
Sysmex XF-1600	1

# Performance monitoring

- With bead reference material, 100%

<b>Bead material</b>	<b>Number of laboratories</b>
BD Biosciences, cytometer Setup and Tracking beads (CST beads)	29
Beckman-Coulter Flow-Check Fluorospheres	7
Beckman-Coulter Flow-Check Pro Fluorospheres	5
BD Biosciences 7-color setup beads	1
Beckman-Coulter Flow-Set Fluorospheres	1
Not mentioned	6

- Additional cellular material, 80%
  - stabilized cells with target ranges for common cell subsets

# Methodology

- T/B/NK assay: 55% lyse no wash

<i>Lysing reagent</i>	Number of laboratories				
BD Biosciences FACS Lysing Solution	25				
Beckman-Coulter VersaLyse	11				
Ammonium chloride (NH <sub>4</sub> Cl)	5				
Beckman-Coulter Optilyse C	3				
BD Biosciences Pharm Lyse	3				
Beckman-Coulter Immunoprep reagent system	1				

	Number of participants				
	CD3 <sup>+</sup>	CD4 <sup>+</sup>	CD8 <sup>+</sup>	CD19 <sup>+</sup>	NK
6 colours	23	23	23	23	23
7 colours	10	10	10	10	10
8 colours	12	12	12	12	12
10 colours	3	3	3	3	3

- $\kappa/\lambda$  assay: 100% labs used  $\geq 2$  washing steps

# Results (CV%)

	% lymphs	# 10 <sup>9</sup> /L
CD3	<5	5-10
CD4	~5	~10
CD8	~5	~10
CD19	10-15	10-15
NK	10-15	10-15
Kappa %B lymph	5-10	NA
Lambda %B lymph	5-10	NA

## Technical validation:

- lymphosum (sum of CD3+% plus CD19+% plus CD3-CD16+ and/or CD56+%) should equal the total lymphocytes gate  $\pm$  5%
- sum of kappa and lambda (expressed as a % of CD19+ B-cells) should be between 90 and 110

# Methodological issues affecting performance










Parameter	Number of evaluated results		% results >3 SD	
	Double platform	Single platform	Double platform	Single platform
CD3 10 <sup>9</sup> /L	268	28	6%	25%
CD4 10 <sup>9</sup> /L	268	28	6%	18%
CD8 10 <sup>9</sup> /L	268	28	5%	18%
CD19 10 <sup>9</sup> /L	268	28	7%	14%
NK cells 10 <sup>9</sup> /L	268	28	7%	14%

Parameter	Number of evaluated results		% results >3 SD	
	Lyse and wash	Lyse no wash	Lyse and wash	Lyse no wash
CD3 %	142	158	6%	4%
CD3 10 <sup>9</sup> /L	138	158	8%	8%
CD4 %	142	158	10%	3%
CD4 10 <sup>9</sup> /L	138	158	9%	6%
CD8 %	142	158	8%	2%
CD8 10 <sup>9</sup> /L	138	158	8%	5%
CD19 %	142	158	8%	4%
CD19 10 <sup>9</sup> /L	138	158	9%	6%
NK cells %	142	158	6%	4%
NK cells 10 <sup>9</sup> /L	138	158	7%	8%
Lymphosum	142	158	5%	6%

# CD34+ cell numeration

Annual report 2024

# ISHAGE-derived single platform protocol



# Methodology

- 75% single platform

<b>Flow cytometer</b>	<b>Number of laboratories</b>
BD Biosciences FACSLyric	9
Beckman-Coulter Navios	4
BD Biosciences FACSCanto II	4
Beckman Coulter AQUIOS CL	2
Miltenyi Biotec MACSQuant analyzer	1

- 14/20: ISHAGE protocol
- 4 BD stem cell enumeration kit
- 1 BD procount
- 1 stem kit Coulter

# Results (%CV), 2024

Sample	Median % CD34+ cells within total WBC	CV %	N	Median CD34+ cells/ $\mu$ L	CV %	N
FC/20466	0.160	12.7	20	9.9	13.5	20
FC/20467	0.640	8.1	20	39.7	11.5	20
FC/20615	0.130	9.4	20	8.3	9.8	20
FC/20616	0.489	5.1	20	30.1	4.4	20

10 cells range: CV ~ 11,65% (median UK NEQAS 12%)

35 cells range: CV ~ 7,95% (median UK NEQAS 8%)

# Total error in DP versus SP CD34+ cell enumeration (UK NEQAS)

%aTE = [Absolute Value (Reported Result–Median)/Median] × 100

		Median CD34+ cells/μL	(Exercises, n) entries, n	Median %aTE (interquartiles)
Dual platform 2004–2014	A	0–10.0	(19) 1,712	12.47 % (5.91–22.50)
	B	10.1–25.0	(37) 3,258	9.81 % (4.54–18.78)
	C	25.1–50.0	(29) 2,057	8.31 % (3.70–15.03)
	D	>50.1	(35) 2,429	8.15 % (3.74–15.29)
Single platform 2004–2014	A	0–10.0	(19) 2,820	11.57 % (5.26–21.97)
	B	10.1–25.0	(37) 4,857	8.94 % (4.12–16.91)
	C	25.1–50.0	(29) 5,193	7.50 % (3.46–13.78)
	D	>50.1	(35) 6,196	6.77 % (2.96–12.26)

In the 10-25 CD34+ cells/μL range:  
delta aTE |DP-SP| = 0,87 % i.e. <0,1-0,25 cell/μL

# Where does EQA fit in a 15189 QA program?

1. Performance monitoring
2. Method validation (UM)


Performance monitoring

# EQA: Requirements of ISO 15189 §7.3.7.3


a) Performance monitoring by comparison with other laboratories

- EQA programs for (all) tests and interpretation of results

Absolute Values (cells/ $\mu$ L)



Percentage Values (%)



EQA yearly planning (e.g., technicians)



data record with full traceability (instrument, calibration, IQC, reagent lot, operator)



timely result review



OK



archive



KO



Corrective action



Efficiency assessment



Full communication!!



# Non-Conformité

Id 105339 Date 31/07/2025 Enregistré par GOTHOT André

Lié à [105338] [QCE.PRO - 3 - Rapport QCE]

Date réelle ou date de détection du problème 31/07/2025 Mot(s)-Clé(s) CD34 - 2025 - 2 TEST

Secteur concerné BCL - HIH - H - Typage hématopoïétique et Thérapie cellulaire

Diffuser la NC au laboratoire

**Enregistrement** Documents Analyse Actions correctives Evaluation de l'efficacité Historique Traçabilité

## Observation

KO  
FC21357: z score > 3  
overestimation by wrong CD45 gating  
FC21358: OK

Fournisseur concerné Sciensano

Sous-Traitant concerné

Est-ce que ce problème est déjà survenu ? Oui

## La NC a un impact sur

ISO 15189	ISO 17025	Item
7.3.7		Garantie de la validité des résultats d'examen(s)

## Origine de la NC

Laboratoire / Secteur commun  
QCE

## Actions immédiates prises







check IQC and patient results  
check competency monitoring of TN and CS  
review CD34 procedure with TN and CS (sample prep, analysis)  
communication to all techs

La Non-Conformité nécessite une analyse des causes profondes ?

Oui

Non

Worklist\_277 110723033301  
TRANS6939 SMS0 0007 8845 302,  
19981224TRANS693\_\_02 34  
20250723 151927  
[Ungated]



# TIP et Communications

BCL - HIH - H - Typage Hématopoïétique et Thérapie Cellulaire

Id 52298 Date 15/11/2019

Projets en cours Projets clôturés Tableaux blancs et Communications Personnel Validation Analyses Suivi annuel TAT Maîtrise des risques

Année Enreg

Id	Date	Description	Conte
<b>Année Enreg : 2024</b>			
96344	26/07/2024	CORRESPONDANT QC	A l'en
95854	10/07/2024	MRD -> PAS de LW	Pour l
95654	26/06/2024	!!!CARRY OVER!!!	2 con
95122	26/06/2024	TRBC1 - CD16	Nouv
94478	02/05/2024	CHIM33	Attent
93610	28/04/2024	FORMOL - HOTTE	Accès
91273	28/03/2024	CD34 (audit FACT	Suite
<b>Année Enreg : 2023</b>			
<b>Année Enreg : 2022</b>			
<b>Année Enreg : 2021</b>			
<b>Année Enreg : 2020</b>			
<b>Année Enreg : 2019</b>			

Sauvegarder et clôturer l'étape

Etape actuelle: 1. Enregistrement

## TIP - Communication

Id < nouve Date 01/08/2025 Enregistré par GOTHOT André

Secteur BCL - HIH - H - Typage Hématopoïétique et Thérapie Cellulaire

Si TB, peut-il être considéré comme une formation continue ?  
 Oui  Non

Date ap. J.-C. Mot(s)-clé(s)

Type de communication QCE

Durée en heure (0,25 = 15 min.)

### Contenu de la communication

error EQA Sciensano CD34/2025/02, review gating procedure

### Commentaires et/ou Suivi des TAT

### Document(s) Traçabilité Présences

Date	Description	Extension	Er
01/08/2025	CD34 ISHAGE, Rapport	.pdf	Gi
01/08/2025	printViewReport (13)	.pdf	Gi

EQA and method validation

# EQA and method validation ISO 15189 7.3.4

- 7.3.4 Measurement uncertainty evaluation (MU)

MU for measured quantitative values shall be evaluated and maintained for its intended use, if relevant. MU shall be compared to performance specifications and documented.

$$U(c) = \sqrt{u^2(IQC) + u^2(K_{EQC})}$$

- $u(IQC)$  = between-run imprecision, stdev IQC
- $u(K_{EQC})$  = bias, max bias / $\sqrt{3}$

## BIAS EQC

% CD34			
	result	median	bias
2023/1	0,556	0,58	-0,024
2023/2	0,55	0,534	0,016
2023/3	0,569	0,593	-0,024
2024/1	0,574	0,595	-0,021
2024/2	0,54	0,64	-0,1
2024/3	0,487	0,489	-0,002
2025/1	0,442	0,477	-0,035
2025/2	0,488	0,488	0
MAX			-0,035
$u(K_{EQC})$			0,020
$u^2(K_{EQC})$			0,0004

## IMPRECISION IQC

Run	CD45 c.	CD34 c.	%CD34
1	153.670	869	0,5655
2	139.824	770	0,5507
3	154.144	861	0,5586
4	155.935	891	0,5714
5	156.621	894	0,5708
6	156.052	899	0,5761
7	153.162	916	0,5981
8	154.455	934	0,6047
9	154.633	888	0,5743
mean			0,5745
$u(IQC)$			0,0173
$u^2(IQC)$			0,00030

$$U(c) = \sqrt{u^2(IQC) + u^2(K_{EQC})}$$

$$U(c) = \sqrt{0,0004 + 0,00030}$$

U	0,0266
$U_{95\%}(k=2)$	0,0532

Absolute error	$0,5745 \pm 0,0532$
Relative error	$0,5745 \pm 9,23\%$

# Additional EQA programs - flow cytometry

- UK NEQAS – INSTAND (DE)
  - CD34+ stem cell enumeration
  - Leukemia/lymphoma immunophenotyping
  - Paroxysmal nocturnal hemoglobinuria
  - Measurable residual disease (MRD): ALL, AML, CLL, MM (UK)
  - CSF immunophenotyping (UK)
- Euroflow
  - Lymphocyte screening tube (LST) (w\*)
  - MRD ALL (w\*/d)
  - MRD MM (d)
  - Primary immunodeficiency orientation tube (PIDOT) (w\*/d)
    - \*Wet lab includes standardisation of fluorescence intensities
    - Digital: analysis of raw data



THANKS FOR YOUR ATTENTION

**CHU**  
de Liège

 **Onilab Lg.**